



Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

August 18, 2011

Sandy Wackett
DEQ LABORATORY
602 N. Fifth St.
Baton Rouge, LA 70802

RE: Project 20127059
Project ID: WEST PEARL RIVER/TEMPLE INLAND

Dear Sandy Wackett:

Enclosed are the analytical results for sample(s) received by the laboratory on August 16, 2011. Results reported herein conform to the most current NELAC standards, where applicable, unless otherwise narrated in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in cursive script that reads "Cindy Olavesen".

Cindy Olavesen
cindy.olavesen@pacelabs.com



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, Inc.

Cover No Results 8/18/2011 11:59

Project: 20127059

Client: DEQ LABORATORY

Project ID: WEST PEARL RIVER/TEMPLE INLAND

Washington Department of Ecology C2078
Oregon Environmental Laboratory Accreditation - LA200001
U.S. Dept. of Agriculture Foreign Soil Import P330-10-00119
Pennsylvania Dept. of Env Protection (NELAC) 68-04202
Texas Commission on Env. Quality (NELAC) T104704405-09-TX
Kansas Department of Health and Environment (NELAC) E-10266
Florida Department of Health (NELAC) E87595
Oklahoma Department of Environmental Quality - 2010-139
Illinois Environmental Protection Agency - 0025721
California Env. Lab Accreditation Program Branch - 11277CA
Louisiana Dept. of Environmental Quality (NELAC/LELAP) 02006

Preliminary Data

8/18/2011 11:59:16



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, Inc.



Sample Cross Reference

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Project: 20127059

Client: DEQ LABORATORY

Project ID: WEST PEARL RIVER/TEMPLE INLAND

Client Sample ID	Lab ID	Matrix	Collection Date/Time	Received Date/Time
506-110816-01	20911089	Water	16-Aug-11 12:35	16-Aug-11 16:00
506-110816-02	20911090	Water	16-Aug-11 13:45	16-Aug-11 16:00

Preliminary Data



Sample Results

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Client: DEQ LABORATORY

Client ID: 506-110816-01

Project: 20127059

Project ID: WEST PEARL RIVER/TEMPLE INLAND

Site: None

Lab ID: 20911089

Matrix: Water

% Moisture: n/a

Description: None

Prep Level: Water

Batch: 166978

Method: EPA 624

GCMS VOAs Trace Water

Collected: 16-Aug-11

Received: 16-Aug-11

Prepared: 17-Aug-11

Units: ug/L

CAS No.	Analyte	Dilution	Result	Qu	Reporting Limit	Reg Limit	Analysis
71-43-2	Benzene	1	ND		0.500		17-Aug-11 14:27 MLS
75-27-4	Bromodichloromethane	1	ND		0.500		17-Aug-11 14:27 MLS
75-25-2	Bromoform	1	ND		0.500		17-Aug-11 14:27 MLS
74-83-9	Bromomethane	1	ND		0.500		17-Aug-11 14:27 MLS
56-23-5	Carbon tetrachloride	1	ND		0.500		17-Aug-11 14:27 MLS
108-90-7	Chlorobenzene	1	ND		0.500		17-Aug-11 14:27 MLS
75-00-3	Chloroethane	1	ND		0.500		17-Aug-11 14:27 MLS
67-66-3	Chloroform	1	ND		0.500		17-Aug-11 14:27 MLS
74-87-3	Chloromethane	1	ND		0.500		17-Aug-11 14:27 MLS
124-48-1	Dibromochloromethane	1	ND		0.500		17-Aug-11 14:27 MLS
95-50-1	1,2-Dichlorobenzene	1	ND		0.500		17-Aug-11 14:27 MLS
541-73-1	1,3-Dichlorobenzene	1	ND		0.500		17-Aug-11 14:27 MLS
106-46-7	1,4-Dichlorobenzene	1	ND		0.500		17-Aug-11 14:27 MLS
75-34-3	1,1-Dichloroethane	1	ND		0.500		17-Aug-11 14:27 MLS
107-06-2	1,2-Dichloroethane	1	ND		0.500		17-Aug-11 14:27 MLS
75-35-4	1,1-Dichloroethene	1	ND		0.500		17-Aug-11 14:27 MLS
156-60-5	trans-1,2-Dichloroethene	1	ND		0.500		17-Aug-11 14:27 MLS
78-87-5	1,2-Dichloropropane	1	ND		0.500		17-Aug-11 14:27 MLS
10061-01-5	cis-1,3-Dichloropropene	1	ND		0.500		17-Aug-11 14:27 MLS
10061-02-6	trans-1,3-Dichloropropene	1	ND		0.500		17-Aug-11 14:27 MLS
100-41-4	Ethylbenzene	1	ND		0.500		17-Aug-11 14:27 MLS
75-09-2	Methylene chloride	1	ND		0.500		17-Aug-11 14:27 MLS
1634-04-4	Methyl-tert-butyl ether	1	ND		0.500		17-Aug-11 14:27 MLS
79-34-5	1,1,2,2-Tetrachloroethane	1	ND		0.500		17-Aug-11 14:27 MLS
127-18-4	Tetrachloroethene	1	ND		0.500		17-Aug-11 14:27 MLS
108-88-3	Toluene	1	ND		0.500		17-Aug-11 14:27 MLS
71-55-6	1,1,1-Trichloroethane	1	ND		0.500		17-Aug-11 14:27 MLS
79-00-5	1,1,2-Trichloroethane	1	ND		0.500		17-Aug-11 14:27 MLS
79-01-6	Trichloroethene	1	ND		0.500		17-Aug-11 14:27 MLS
75-69-4	Trichlorofluoromethane	1	ND		0.500		17-Aug-11 14:27 MLS
75-01-4	Vinyl chloride	1	ND		0.500		17-Aug-11 14:27 MLS

31 compound(s) reported

ND denotes the analyte was analyzed for but not detected at the reporting limit or method detection limit indicated.

Protocol 8/18/2011 11:59:46
Limits are corrected for sample size, dilution and moisture content if applicable.

Qu lists qualifiers. Specific qualifiers are defined at the end of the report.

Regulatory limit may denote an actual regulatory limit or a client-requested notification limit.



Sample Results

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Client: DEQ LABORATORY

Client ID: 506-110816-02

Project: 20127059

Project ID: WEST PEARL RIVER/TEMPLE INLAND

Site: None

Lab ID: 20911090

Matrix: Water

% Moisture: n/a

Description: None

Prep Level: Water

Batch: 166978

Method: EPA 624

GCMS VOAs Trace Water

Collected: 16-Aug-11

Received: 16-Aug-11

Prepared: 17-Aug-11

Units: ug/L

CAS No.	Analyte	Dilution	Result	Qu	Reporting Limit	Reg Limit	Analysis
71-43-2	Benzene	1	ND		0.500		17-Aug-11 14:49 MLS
75-27-4	Bromodichloromethane	1	ND		0.500		17-Aug-11 14:49 MLS
75-25-2	Bromoform	1	ND		0.500		17-Aug-11 14:49 MLS
74-83-9	Bromomethane	1	ND		0.500		17-Aug-11 14:49 MLS
56-23-5	Carbon tetrachloride	1	ND		0.500		17-Aug-11 14:49 MLS
108-90-7	Chlorobenzene	1	ND		0.500		17-Aug-11 14:49 MLS
75-00-3	Chloroethane	1	ND		0.500		17-Aug-11 14:49 MLS
67-66-3	Chloroform	1	ND		0.500		17-Aug-11 14:49 MLS
74-87-3	Chloromethane	1	ND		0.500		17-Aug-11 14:49 MLS
124-48-1	Dibromochloromethane	1	ND		0.500		17-Aug-11 14:49 MLS
95-50-1	1,2-Dichlorobenzene	1	ND		0.500		17-Aug-11 14:49 MLS
541-73-1	1,3-Dichlorobenzene	1	ND		0.500		17-Aug-11 14:49 MLS
106-46-7	1,4-Dichlorobenzene	1	ND		0.500		17-Aug-11 14:49 MLS
75-34-3	1,1-Dichloroethane	1	ND		0.500		17-Aug-11 14:49 MLS
107-06-2	1,2-Dichloroethane	1	ND		0.500		17-Aug-11 14:49 MLS
75-35-4	1,1-Dichloroethene	1	ND		0.500		17-Aug-11 14:49 MLS
156-60-5	trans-1,2-Dichloroethene	1	ND		0.500		17-Aug-11 14:49 MLS
78-87-5	1,2-Dichloropropane	1	ND		0.500		17-Aug-11 14:49 MLS
10061-01-5	cis-1,3-Dichloropropene	1	ND		0.500		17-Aug-11 14:49 MLS
10061-02-6	trans-1,3-Dichloropropene	1	ND		0.500		17-Aug-11 14:49 MLS
100-41-4	Ethylbenzene	1	ND		0.500		17-Aug-11 14:49 MLS
75-09-2	Methylene chloride	1	ND		0.500		17-Aug-11 14:49 MLS
1634-04-4	Methyl-tert-butyl ether	1	ND		0.500		17-Aug-11 14:49 MLS
79-34-5	1,1,2,2-Tetrachloroethane	1	ND		0.500		17-Aug-11 14:49 MLS
127-18-4	Tetrachloroethene	1	ND		0.500		17-Aug-11 14:49 MLS
108-88-3	Toluene	1	ND		0.500		17-Aug-11 14:49 MLS
71-55-6	1,1,1-Trichloroethane	1	ND		0.500		17-Aug-11 14:49 MLS
79-00-5	1,1,2-Trichloroethane	1	ND		0.500		17-Aug-11 14:49 MLS
79-01-6	Trichloroethene	1	ND		0.500		17-Aug-11 14:49 MLS
75-69-4	Trichlorofluoromethane	1	ND		0.500		17-Aug-11 14:49 MLS
75-01-4	Vinyl chloride	1	ND		0.500		17-Aug-11 14:49 MLS

31 compound(s) reported

ND denotes the analyte was analyzed for but not detected at the reporting limit or method detection limit indicated.

Protocol 8/18/2011 11:59:46
Limits are corrected for sample size, dilution and moisture content if applicable.

Qu lists qualifiers. Specific qualifiers are defined at the end of the report.

Regulatory limit may denote an actual regulatory limit or a client-requested notification limit.



Sample Results

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Client: DEQ LABORATORY

Client ID: 506-110816-01

Project: 20127059

Project ID: WEST PEARL RIVER/TEMPLE INLAND

Site: None

Lab ID: 20911089

Matrix: Water

%Moisture: n/a

Description: None

Collected: 16-Aug-11

Received: 16-Aug-11

Analyte	Method	Batch	DF	Result	Qu	Units	Reporting Limit	Prep.	Analysis	Limit
Oil and Grease	EPA 1664A	166995	1	ND		mg/L	5.00	17-Aug-11	17-Aug-11 13:27	JMA
1 parameter(s) reported										

Preliminary Data

ND denotes the analyte was analyzed for but not detected at the reporting limit or method detection limit indicated.

Inorganics 8/18/2011 11:59:47

Limits are corrected for sample size, dilution and moisture content if applicable.

Qu lists qualifiers. Specific qualifiers are defined at the end of the report.

Regulatory limit may denote an actual regulatory limit or a client-requested notification limit.

Batch: 166978

Project: 20127059

Method: Water GC/MS Volatile Organics

Lab ID	Sample ID	Qu	Sur 1 %Rec	Sur 2 %Rec	Sur 3 %Rec	Sur 4 %Rec	Sur 5 %Rec	Sur 6 %Rec	Sur 7 %Rec	Sur 8 %Rec
20911290	166978 BLANK 1		94	77	89					
20911291	166978 LCS 1		94	77	89					
20911089	506-110816-01		94	78	89					
20911292	506-110816-01 MS 1		92	75	86					
20911293	506-110816-01 MSD 1		92	75	88					
20911090	506-110816-02		95	77	89					

QC limits:

68-124

72-126

79-119

Sur 1: 4-Bromofluorobenzene (S)

Sur 2: Dibromofluoromethane (S)

Sur 3: Toluene-d8 (S)

Preliminary Data



Quality Control

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Batch: 166978
Method: Water GC/MS Volatile Organics

Project: 20127059 **LCS:** 20911291 17-Aug-11 13:00
MS: 20911292 17-Aug-11 13:22
Units: ug/L **MSD:** 20911293 17-Aug-11 13:43
Original for MS: Client Sample 20911089

Parameter Name	LCS Spike	LCS Found	LCS %Rec	MS Spike	Sample Found	MS Found	MSD Found	MS %Rec	MSD %Rec	RPD	QC Limits		Max	Qu
											LCS	MS/MSD	RPD	
Benzene	50.0	54.6	109	50.0		55.7	54.1	112	108	3	62-131	52-141	20	
Bromodichloromethane	50.0	54.4	109	50.0		54.8	53.8	110	108	2	69-132	70-134	20	
Bromoform	50.0	58.5	117	50.0		57.7	58.0	115	116	1	35-166	37-171	20	
Bromomethane	50.0	41.8	84	50.0		40.1	38.8	80	78	3	34-158	34-155	20	
Carbon tetrachloride	50.0	52.8	106	50.0		53.9	51.5	108	103	5	54-144	48-146	20	
Chlorobenzene	50.0	58.3	117	50.0		59.7	57.6	119	115	4	70-127	67-129	20	
Chloroethane	50.0	42.2	84	50.0		43.1	41.0	86	82	5	17-195	12-192	20	
Chloroform	50.0	48.5	97	50.0		49.3	48.1	99	96	3	73-134	66-143	20	
Chloromethane	50.0	39.1	78	50.0		39.3	36.6	79	73	7	17-153	14-155	20	
Dibromochloromethane	50.0	57.6	115	50.0		58.2	57.3	116	115	2	49-146	50-150	20	
1,2-Dichlorobenzene	50.0	57.7	115	50.0		57.5	56.6	115	113	2	58-130	62-131	20	
1,3-Dichlorobenzene	50.0	56.7	113	50.0		57.3	55.9	115	112	3	60-125	58-128	20	
1,4-Dichlorobenzene	50.0	56.0	112	50.0		57.0	54.6	114	109	4	60-128	59-130	20	
1,1-Dichloroethane	50.0	48.2	96	50.0		49.7	47.6	99	95	4	63-129	59-133	20	
1,2-Dichloroethane	50.0	54.5	109	50.0		54.4	54.1	109	108	1	57-148	56-154	20	
1,1-Dichloroethene	50.0	43.6	87	50.0		46.1	43.2	92	86	7	51-139	44-146	20	
trans-1,2-Dichloroethene	50.0	44.6	89	50.0		46.9	44.1	94	88	6	60-129	57-132	20	
1,2-Dichloropropane	50.0	54.5	109	50.0		55.3	53.9	111	108	3	66-128	62-135	20	
cis-1,3-Dichloropropene	50.0	54.6	109	50.0		54.9	54.2	110	108	1	72-138	70-139	20	
trans-1,3-Dichloropropene	50.0	54.9	110	50.0		55.3	54.9	111	110	1	59-149	59-154	20	
Ethylbenzene	50.0	57.0	114	50.0		58.7	55.9	117	112	5	66-126	57-135	20	
Methylene chloride	50.0	45.5	91	50.0		43.1	41.9	86	84	3	46-168	45-166	20	
Methyl-tert-butyl ether	50.0	49.5	99	50.0		49.8	49.8	100	100	0	37-166	35-176	20	
1,1,2,2-Tetrachloroethane	50.0	54.2	108	50.0		52.9	52.8	106	106	0	15-179	15-187	20	
Tetrachloroethene	50.0	61.0	122	50.0		63.6	60.6	127	121	5	46-157	48-143	20	
Toluene	50.0	55.7	111	50.0		56.6	55.0	113	110	3	69-126	59-136	20	
1,1,1-Trichloroethane	50.0	49.8	100	50.0		51.0	49.1	102	98	4	62-131	54-137	20	
1,1,2-Trichloroethane	50.0	55.5	111	50.0		56.1	56.2	112	112	0	58-144	59-148	20	
Trichloroethene	50.0	57.0	114	50.0		58.5	56.2	117	113	4	67-132	58-140	20	
Trichlorofluoromethane	50.0	45.8	92	50.0		46.6	44.3	93	89	5	39-171	24-175	20	
Vinyl chloride	50.0	44.3	89	50.0		45.8	42.8	92	86	7	27-149	21-150	20	

31 compound(s) reported



Blank Results

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Blank ID: 166978 BLANK 1

Project: 20127059

Lab ID: 20911290

Prep Level: Water

Batch: 166978

Method: Water GC/MS Volatile Organics

Prepared: 17-Aug-11

						Units: <u>ug/L</u>		
CAS Numb	Analyte	Dilution	Result	Qu	Reporting Limit		Analysis	
71-43-2	Benzene	1	ND		0.500		17-Aug-11 12:38	MLS
75-27-4	Bromodichloromethane	1	ND		0.500		17-Aug-11 12:38	MLS
75-25-2	Bromoform	1	ND		0.500		17-Aug-11 12:38	MLS
74-83-9	Bromomethane	1	ND		0.500		17-Aug-11 12:38	MLS
56-23-5	Carbon tetrachloride	1	ND		0.500		17-Aug-11 12:38	MLS
108-90-7	Chlorobenzene	1	ND		0.500		17-Aug-11 12:38	MLS
75-00-3	Chloroethane	1	ND		0.500		17-Aug-11 12:38	MLS
67-66-3	Chloroform	1	ND		0.500		17-Aug-11 12:38	MLS
74-87-3	Chloromethane	1	ND		0.500		17-Aug-11 12:38	MLS
124-48-1	Dibromochloromethane	1	ND		0.500		17-Aug-11 12:38	MLS
95-50-1	1,2-Dichlorobenzene	1	ND		0.500		17-Aug-11 12:38	MLS
541-73-1	1,3-Dichlorobenzene	1	ND		0.500		17-Aug-11 12:38	MLS
106-46-7	1,4-Dichlorobenzene	1	ND		0.500		17-Aug-11 12:38	MLS
75-34-3	1,1-Dichloroethane	1	ND		0.500		17-Aug-11 12:38	MLS
107-06-2	1,2-Dichloroethane	1	ND		0.500		17-Aug-11 12:38	MLS
75-35-4	1,1-Dichloroethene	1	ND		0.500		17-Aug-11 12:38	MLS
156-60-5	trans-1,2-Dichloroethene	1	ND		0.500		17-Aug-11 12:38	MLS
78-87-5	1,2-Dichloropropane	1	ND		0.500		17-Aug-11 12:38	MLS
10061-01-5	cis-1,3-Dichloropropene	1	ND		0.500		17-Aug-11 12:38	MLS
10061-02-6	trans-1,3-Dichloropropene	1	ND		0.500		17-Aug-11 12:38	MLS
100-41-4	Ethylbenzene	1	ND		0.500		17-Aug-11 12:38	MLS
75-09-2	Methylene chloride	1	ND		0.500		17-Aug-11 12:38	MLS
1634-04-4	Methyl-tert-butyl ether	1	ND		0.500		17-Aug-11 12:38	MLS
79-34-5	1,1,2,2-Tetrachloroethane	1	ND		0.500		17-Aug-11 12:38	MLS
127-18-4	Tetrachloroethene	1	ND		0.500		17-Aug-11 12:38	MLS
108-88-3	Toluene	1	ND		0.500		17-Aug-11 12:38	MLS
71-55-6	1,1,1-Trichloroethane	1	ND		0.500		17-Aug-11 12:38	MLS
79-00-5	1,1,2-Trichloroethane	1	ND		0.500		17-Aug-11 12:38	MLS
79-01-6	Trichloroethene	1	ND		0.500		17-Aug-11 12:38	MLS
75-69-4	Trichlorofluoromethane	1	ND		0.500		17-Aug-11 12:38	MLS
75-01-4	Vinyl chloride	1	ND		0.500		17-Aug-11 12:38	MLS

31 compound(s) reported

ND denotes the analyte was analyzed for but not detected at the reporting limit or method detection limit indicated.

Protocol Blank 8/18/2011 11:59:5
Limits are corrected for sample size, dilution and moisture content if applicable.

Qu lists qualifiers. Specific qualifiers are defined at the end of the report.

Regulatory limit may denote an actual regulatory limit or a client-requested notification limit.



Inorganics Quality Control

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Project: 20127059

Parameter	Batch	Blank	Units	LCS Spike	LCS Found	LCS %Rec	MS Spike	Sample Found	MS Found	MSD Found	MS %Rec	MSD %Rec	RPD	QC Limits		Max	Qu
Oil and Grease	166995	< 5.00	mg/L	40.0	33.5	84	42.1		34.4		80			78-114	78-114		

Preliminary Data

* denotes recovery outside of QC limits.

MS/MSD RPD is calculated via SW-846 rules on the basis of spiked sample concentrations rather than spike recoveries.



Definitions/Qualifiers

Pace Analytical Services, Inc.
1000 Riverbend Blvd. Suite F
St. Rose, LA 70087
(504) 469-0333

Project: 20127059

Value	Description
J	This estimated value for the analyte is below the adjusted reporting limit but above the instrument reporting limit.
U	The analyte was analyzed for but not detected at the reporting limit or method detection limit indicated.
B	This analyte was detected in the method blank.
E	The sample concentration is above the linear calibrated range of the analysis.
LCS	Laboratory Control Sample.
MS(D)	Matrix Spike (Duplicate).
DUP	Sample Duplicate.
RPD	Relative Percent Difference.

Preliminary Data

Chains of Custody

Preliminary Data



Sample Condition

1000 Riverbend Blvd., Suite F
St. Rose, LA 70087Courier: ☐ Pace Courier ☐ Hackbarth ☐ Fed X ☐ UPS ☐ DHL ☐ USPS ☒ Customer ☐ Other

Custody Seal on Cooler/Box Present: [see COC]

Custody Seals Intact: ☐ Yes ☒ No

Thermometer

Used:

- ☒ Therm Fisher IR 1
☐ Therm Fisher IR 2
☐ Therm Fisher IR 4

Type of Ice: Wet Blue None

Samples on ice: [see COC]

Cooler Temperature: [see COC]

Temp should be above freezing to 6°C

Date and Initials of person examining
contents: 8-16-12

Temp must be measured from Temperature blank when present

Comments:

Temperature Blank Present?	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	1
Chain of Custody Present:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	2
Chain of Custody Complete:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	3
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4
Sampler Name & Signature on COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	5
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	6
Sufficient Volume:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	7
Correct Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	8
Filtered vol. Rec. for Diss. tests	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	9
Sample Labels match COC:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	10
All containers received within manufacture's precautionary and/or expiration dates.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	11
All containers needing preservation have been checked (except VOA, coliform, & O&G).	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	12
All containers preservation checked found to be in compliance with EPA recommendation.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13
	If No, was preservative added? <input type="checkbox"/> Yes <input type="checkbox"/> No If added record lot no.: HNO3 _____ H2SO4 _____	
Samples checked for dechlorination:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> N/A	14
Trip Blank Present:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17
Pace Trip Blank Lot # (if purchased):	<u>N/A</u>	18

Client Notification/ Resolution:

Person Contacted: _____

Date/Time: _____

Comments/ Resolution: _____